

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC040821AL\
 Data File : FC053017.D
 Signal(s) : FID1A.CH
 Acq On : 08 Apr 2021 21:00
 Operator : DD\AJ
 Sample : 20 PPM ALIPHATIC STD ICV
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 20 PPM ALIPHATIC STD ICV

Integration File: autoint1.e
 Quant Time: Apr 09 09:26:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 040821.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 09 09:23:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	125.160	117.954 E3	5.8	94	0.00
2 T n-Decane (C10)	125.893	119.054 E3	5.4	95	0.00
3 T A~Naphthalene (C11.7)	139.299	131.567 E3	5.6	94	0.00
4 T n-Dodecane (C12)	128.074	120.967 E3	5.5	95	0.00
5 T A~2-methylnaphthalene (C12.	141.358	133.673 E3	5.4	94	0.00
6 T n-Tetradecane (C14)	128.667	122.027 E3	5.2	95	0.00
7 T n-Hexadecane (C16)	127.857	120.991 E3	5.4	95	0.00
8 T n-Octadecane (C18)	135.240	127.925 E3	5.4	95	0.00
9 S ortho-Terphenyl (SURR)	144.471	136.859 E3	5.3	95	0.00
10 T n-Eicosane (C20)	137.355	130.113 E3	5.3	95	0.00
11 T n-Heneicosane (C21)	137.402	130.529 E3	5.0	95	0.00
12 S 1-chlorooctadecane (SURR)	115.542	110.183 E3	4.6	95	0.00
13 T n-Docosane (C22)	137.839	131.004 E3	5.0	95	0.00
14 T n-Tetracosane (C24)	137.879	131.392 E3	4.7	95	0.00
15 T n-Hexacosane (C26)	135.603	129.719 E3	4.3	95	0.00
16 T n-Octacosane (C28)	133.294	127.420 E3	4.4	95	0.00
17 T n-Tricontane (C30)	128.339	123.045 E3	4.1	96	0.00
18 T n-Dotriacontane (C32)	119.470	114.522 E3	4.1	95	0.00
19 T n-Tetratriacontane (C34)	110.751	106.247 E3	4.1	95	0.00
20 T n-Hexatriacontane (C36)	100.910	96.280 E3	4.6	94	0.00
21 T n-Octatriacontane (C38)	89.091	86.094 E3	3.4	96	0.00
22 T n-Tetracontane (C40)	80.054	78.472 E3	2.0	99	0.00

Evaluate Continuing Calibration Report - Not Founds

(#) = Out of Range

SPCC's out = 0 CCC's out = 0