

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC040921AL\
 Data File : FC053038.D
 Signal(s) : FID1A.CH
 Acq On : 09 Apr 2021 11:28
 Operator : DD\AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Apr 10 00:12:49 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 040821.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 09 09:23:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	125.160	126.544 E3	-1.1	101	0.00
2 T n-Decane (C10)	125.893	128.090 E3	-1.7	102	0.00
3 T A~Naphthalene (C11.7)	139.299	142.435 E3	-2.3	102	0.00
4 T n-Dodecane (C12)	128.074	131.700 E3	-2.8	103	0.00
5 T A~2-methylnaphthalene (C12.	141.358	145.343 E3	-2.8	103	0.00
6 T n-Tetradecane (C14)	128.667	132.228 E3	-2.8	103	0.00
7 T n-Hexadecane (C16)	127.857	129.922 E3	-1.6	102	0.00
8 T n-Octadecane (C18)	135.240	134.236 E3	0.7	100	0.00
9 S ortho-Terphenyl (SURR)	144.471	139.801 E3	3.2	97	0.00
10 T n-Eicosane (C20)	137.355	132.599 E3	3.5	97	0.00
11 T n-Heneicosane (C21)	137.402	130.367 E3	5.1	95	0.00
12 S 1-chlorooctadecane (SURR)	115.542	110.540 E3	4.3	96	0.00
13 T n-Docosane (C22)	137.839	128.755 E3	6.6	94	0.00
14 T n-Tetracosane (C24)	137.879	124.031 E3	10.0	90	0.00
15 T n-Hexacosane (C26)	135.603	117.846 E3	13.1	87	0.00
16 T n-Octacosane (C28)	133.294	112.166 E3	15.9	84	0.00
17 T n-Tricontane (C30)	128.339	107.030 E3	16.6	83	0.00
18 T n-Dotriacontane (C32)	119.470	102.813 E3	13.9	86	0.00
19 T n-Tetratriacontane (C34)	110.751	98.620 E3	11.0	89	0.00
20 T n-Hexatriacontane (C36)	100.910	96.519 E3	4.4	94	0.00
21 T n-Octatriacontane (C38)	89.091	92.614 E3	-4.0	103	0.00
22 T n-Tetracontane (C40)	80.054	96.271 E3	-20.3	121	0.01

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0