

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC041223AL\
 Data File : FC063356.D
 Signal(s) : FID1A.ch
 Acq On : 12 Apr 2023 20:02
 Operator : YP/AJ
 Sample : 02257-03
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 ELI-WATER-0407

Integration File: autoint1.e
 Quant Time: Apr 13 01:57:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 040523.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 05 07:22:46 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.974	4593111	41.512 ug/ml
Spiked Amount	50.000	Recovery	= 83.02%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC041223AL\
Data File : FC063356.D
Signal(s) : FID1A.ch
Acq On : 12 Apr 2023 20:02
Operator : YP/AJ
Sample : 02257-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
ELI-WATER-0407

Integration File: autoint1.e
Quant Time: Apr 13 01:57:30 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 040523.M
Quant Title : GC Extractables
QLast Update : Wed Apr 05 07:22:46 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

