

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC041321AL\
 Data File : FC053129.D
 Signal(s) : FID1A.CH
 Acq On : 14 Apr 2021 00:11
 Operator : DD\AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Apr 14 02:15:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 040821.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 09 09:23:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	125.160	122.113 E3	2.4	98	0.00
2 T n-Decane (C10)	125.893	123.817 E3	1.6	98	0.00
3 T A~Naphthalene (C11.7)	139.299	136.060 E3	2.3	98	0.00
4 T n-Dodecane (C12)	128.074	127.517 E3	0.4	100	0.00
5 T A~2-methylnaphthalene (C12.	141.358	139.852 E3	1.1	99	0.00
6 T n-Tetradecane (C14)	128.667	127.230 E3	1.1	99	0.00
7 T n-Hexadecane (C16)	127.857	125.382 E3	1.9	98	0.00
8 T n-Octadecane (C18)	135.240	128.979 E3	4.6	96	0.00
9 S ortho-Terphenyl (SURR)	144.471	134.366 E3	7.0	93	0.00
10 T n-Eicosane (C20)	137.355	127.547 E3	7.1	93	0.00
11 T n-Heneicosane (C21)	137.402	125.392 E3	8.7	91	0.00
12 S 1-chlorooctadecane (SURR)	115.542	106.409 E3	7.9	92	0.00
13 T n-Docosane (C22)	137.839	123.636 E3	10.3	90	0.00
14 T n-Tetracosane (C24)	137.879	119.376 E3	13.4	87	0.00
15 T n-Hexacosane (C26)	135.603	113.971 E3	16.0	84	0.00
16 T n-Octacosane (C28)	133.294	109.143 E3	18.1	82	0.00
17 T n-Tricontane (C30)	128.339	104.861 E3	18.3	81	0.00
18 T n-Dotriacontane (C32)	119.470	100.454 E3	15.9	84	0.00
19 T n-Tetratriacontane (C34)	110.751	97.122 E3	12.3	87	0.00
20 T n-Hexatriacontane (C36)	100.910	95.101 E3	5.8	92	0.00
21 T n-Octatriacontane (C38)	89.091	91.068 E3	-2.2	101	0.00
22 T n-Tetracontane (C40)	80.054	94.445 E3	-18.0	119	0.01

Evaluate Continuing Calibration Report - Not Founds

(#) = Out of Range

SPCC's out = 0 CCC's out = 0