

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC041524AL\  
Data File : FC066207.D  
Signal(s) : FID1A.ch  
Acq On : 16 Apr 2024 00:11  
Operator : YP/AJ  
Sample : I.BLK  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Integration File: autoint1.e  
Quant Time: Apr 16 08:20:59 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 041524.M  
Quant Title : GC Extractables  
QLast Update : Tue Apr 16 08:19:15 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um

| Compound                    | R.T.   | Response | Conc Units  |
|-----------------------------|--------|----------|-------------|
| -----                       |        |          |             |
| System Monitoring Compounds |        |          |             |
| Target Compounds            |        |          |             |
| 22) T n-Tetracontane (C40)  | 21.746 | 121105   | 1.338 ug/ml |
| -----                       |        |          |             |


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC041524AL\  
Data File : FC066207.D  
Signal(s) : FID1A.ch  
Acq On : 16 Apr 2024 00:11  
Operator : YP/AJ  
Sample : I.BLK  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Integration File: autoint1.e  
Quant Time: Apr 16 08:20:59 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 041524.M  
Quant Title : GC Extractables  
QLast Update : Tue Apr 16 08:19:15 2024  
Response via : Initial Calibration  
Integrator: ChemStation


Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURRE)

R.T.: 0.000 min  
Exp R.T.: 11.573 min  
Response: 0  
Conc: N.D.



#12 1-chlorooctadecane (SURRE)

R.T.: 0.000 min  
Exp R.T.: 13.013 min  
Response: 0  
Conc: N.D.



#22 n-Tetracontane (C40)

R.T.: 21.746 min  
Delta R.T.: -0.006 min  
Response: 121105  
Conc: 1.34 ug/ml