

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC041621AL\  
Data File : FC053221.D  
Signal(s) : FID1A.CH  
Acq On : 16 Apr 2021 13:01  
Operator : DD\AJ  
Sample : M2025-03DL 20X  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
FID\_C  
ClientSampleId :

Integration File: autoint1.e  
Quant Time: Apr 16 21:17:27 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Alaphatic EPH 040821.M  
Quant Title : GC Extractables  
QLast Update : Fri Apr 09 09:23:51 2021  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um

| Compound                      | R.T.  | Response | Conc Units  |
|-------------------------------|-------|----------|-------------|
| -----                         |       |          |             |
| System Monitoring Compounds   |       |          |             |
| Target Compounds              |       |          |             |
| 5) T A~2-methylnaphthalene... | 7.018 | 933398   | 6.603 ug/ml |
| -----                         |       |          |             |

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC041621AL\  
Data File : FC053221.D  
Signal(s) : FID1A.CH  
Acq On : 16 Apr 2021 13:01  
Operator : DD\AJ  
Sample : M2025-03DL 20X  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
FID\_C  
ClientSampleId :

Integration File: autoint1.e  
Quant Time: Apr 16 21:17:27 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Alaphatic EPH 040821.M  
Quant Title : GC Extractables  
QLast Update : Fri Apr 09 09:23:51 2021  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.018 min  
Delta R.T.: -0.001 min  
Response: 933398  
Conc: 6.60 ug/ml

Instrument :  
FID\_C  
ClientSampleId :



#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min  
Exp R.T. : 11.551 min  
Response: 0  
Conc: N.D.



#12 1-chlorooctadecane (SURR)

R.T.: 0.000 min  
Exp R.T. : 12.990 min  
Response: 0  
Conc: N.D.