

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC042023AL\
 Data File : FC063444.D
 Signal(s) : FID1A.ch
 Acq On : 21 Apr 2023 12:22
 Operator : YP/AJ
 Sample : 02389-30DL 10X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 SPB-8WC-23-04-17DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/24/2023
 Supervised By :Ankita Jodhani 04/24/2023

Integration File: autoint1.e
 Quant Time: Apr 22 01:35:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 040523.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 05 07:22:46 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.541	297863	1.958 ug/ml
Spiked Amount 50.000		Recovery =	3.92%
12) S 1-chlorooctadecane (S...	12.980	204397	1.847 ug/mlm
Spiked Amount 50.000		Recovery =	3.69%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC042023AL\
Data File : FC063444.D
Signal(s) : FID1A.ch
Acq On : 21 Apr 2023 12:22
Operator : YP/AJ
Sample : 02389-30DL 10X
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
FID_C
ClientSampleId :
SPB-8WC-23-04-17DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/24/2023
Supervised By :Ankita Jodhani 04/24/2023

Integration File: autoint1.e
Quant Time: Apr 22 01:35:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 040523.M
Quant Title : GC Extractables
QLast Update : Wed Apr 05 07:22:46 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.541 min
Delta R.T.: 0.013 min
Response: 297863
Conc: 1.96 ug/ml

Instrument :
FID_C
ClientSampleId :
SPB-8WC-23-04-17DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/24/2023
Supervised By :Ankita Jodhani 04/24/2023



#12 1-chlorooctadecane (SURR)

R.T.: 12.980 min
Delta R.T.: 0.013 min
Response: 204397
Conc: 1.85 ug/ml m