

rters

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/27/2022
Supervised By :mohammad ahmed 04/27/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC04252
Data File : FC059447.D
Signal(s) : FID1A.CH
Acq On : 26 Apr 2022 19:04
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 042122.M
Title : GC Extractables

Signal : FID1A.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.961	2.925	3.065	BB	299733	2657478	82.41%	4.610%
2	4.001	3.958	4.131	BB	275168	2710678	84.06%	4.703%
3	5.551	5.525	5.768	BB	195234	2796139	86.71%	4.851%
4	5.998	5.961	6.121	BB	273270	2866840	88.91%	4.974%
5	6.599	6.511	6.825	BB	204702	2862320	88.77%	4.966%
6	7.781	7.748	7.901	BB	271853	2929313	90.84%	5.082%
7	9.372	9.331	9.481	BB	263540	2991384	92.77%	5.190%
8	10.805	10.765	10.901	BB	258177	3023228	93.76%	5.245%
9	11.097	11.038	11.218	BB	283174	3224551	100.00%	5.594%
10	12.106	12.065	12.198	BB	242426	2968418	92.06%	5.150%
11	12.545	12.501	12.641	BB	203660	2528422	78.41%	4.387%
12	12.715	12.671	12.808	BB	241397	2939624	91.16%	5.100%
13	13.299	13.208	13.385	BB	222626	2922553	90.63%	5.070%
14	14.395	14.311	14.485	BB	209769	2793623	86.64%	4.847%
15	15.411	15.331	15.450	BBA	206449	2640784	81.90%	4.581%
16	16.357	16.278	16.455	BB	186219	2551170	79.12%	4.426%
17	17.239	17.171	17.281	BB	180163	2443033	75.76%	4.238%
18	18.069	17.845	18.135	BB	156785	2321805	72.00%	4.028%
19	18.851	18.801	18.898	BV	138240	2103712	65.24%	3.650%
20	19.587	19.541	19.650	BBA	119632	1899890	58.92%	3.296%
21	20.286	20.251	20.329	BV	115709	1744249	54.09%	3.026%
22	21.028	20.981	21.090	BBA	86930	1721549	53.39%	2.987%
Sum of corrected areas:						57640763		

Aliphatic EPH 042122.M Wed Apr 27 15:31:21 2022