

rteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC050523
Data File : FC063570.D
Signal(s) : FID1A.ch
Acq On : 05 May 2023 20:22
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/08/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 050523.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.300	3.259	3.341	BB	369204	3197415	84.50%	4.912%
2	4.363	4.322	4.410	BB	345801	3246448	85.79%	4.987%
3	5.945	5.909	6.013	BB	344480	3439748	90.90%	5.284%
4	6.380	6.340	6.422	BB	324644	3266549	86.32%	5.018%
5	7.001	6.954	7.080	BB	309687	3304122	87.32%	5.076%
6	8.173	8.126	8.235	BB	294977	3199205	84.55%	4.914%
7	9.771	9.726	9.833	BB	287051	3262016	86.21%	5.011%
8	11.212	11.167	11.271	BB	275476	3276756	86.59%	5.034%
9	11.525	11.476	11.573	BB	348012	3784013	100.00%	5.813%
10	12.517	12.472	12.561	BB	260117	3083219	81.48%	4.736%
11	12.962	12.914	13.006	BB	219718	2661309	70.33%	4.088%
12	13.128	13.086	13.172	BB	237468	2985620	78.90%	4.586%
13	13.714	13.662	13.764	BB	238882	2926771	77.35%	4.496%
14	14.813	14.766	14.863	BB	217978	2837762	74.99%	4.359%
15	15.832	15.794	15.883	BB	200883	2750133	72.68%	4.225%
16	16.779	16.742	16.828	BB	190258	2646104	69.93%	4.065%
17	17.664	17.621	17.714	BB	197369	2667915	70.50%	4.098%
18	18.494	18.444	18.541	BB	174005	2634290	69.62%	4.047%
19	19.275	19.220	19.325	BB	180130	2554699	67.51%	3.924%
20	20.011	19.976	20.045	BB	167696	2435984	64.38%	3.742%
21	20.737	20.677	20.795	BB	147440	2477394	65.47%	3.806%
22	21.627	21.562	21.701	BB	102310	2460837	65.03%	3.780%
Sum of corrected areas:						65098309		

Aliphatic EPH 050523.M Sat May 06 02:18:48 2023