

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC051024AL\
 Data File : FC066375.D
 Signal(s) : FID1A.ch
 Acq On : 10 May 2024 13:50
 Operator : YP/AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: May 11 00:53:56 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 050124.M
 Quant Title : GC Extractables
 QLast Update : Thu May 02 09:03:15 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	151.030	121.360 E3	19.6	73	0.00
2 T n-Decane (C10)	153.784	124.928 E3	18.8	73	0.00
3 T A~Naphthalene (C11.7)	169.896	138.864 E3	18.3	73	0.00
4 T n-Dodecane (C12)	153.682	126.263 E3	17.8	74	0.00
5 T A~2-methylnaphthalene (C12.	164.809	135.117 E3	18.0	73	0.00
6 T n-Tetradecane (C14)	151.338	125.029 E3	17.4	75	0.00
7 T n-Hexadecane (C16)	152.907	128.224 E3	16.1	75	0.00
8 T n-Octadecane (C18)	151.790	128.084 E3	15.6	76	0.00
9 S ortho-Terphenyl (SURR)	165.334	138.438 E3	16.3	75	0.00
10 T n-Eicosane (C20)	143.854	121.796 E3	15.3	77	0.00
11 T n-Heneicosane (C21)	140.567	119.112 E3	15.3	77	0.00
12 S 1-chlorooctadecane (SURR)	122.660	102.947 E3	16.1	76	0.00
13 T n-Docosane (C22)	140.006	118.892 E3	15.1	78	0.00
14 T n-Tetracosane (C24)	140.779	117.298 E3	16.7	78	0.00
15 T n-Hexacosane (C26)	136.784	114.438 E3	16.3	79	0.00
16 T n-Octacosane (C28)	133.920	112.549 E3	16.0	80	0.00
17 T n-Tricontane (C30)	134.763	115.877 E3	14.0	82	0.00
18 T n-Dotriacontane (C32)	128.757	114.594 E3	11.0	85	0.00
19 T n-Tetratriacontane (C34)	118.315	112.199 E3	5.2	88	0.00
20 T n-Hexatriacontane (C36)	114.944	114.424 E3	0.5	85	0.00
21 T n-Octatriacontane (C38)	107.635	121.376 E3	-12.8	104	0.00
22 T n-Tetracontane (C40)	107.233	122.915 E3	-14.6	106	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0