

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC052120AL\  
 Data File : FC047544.D  
 Signal(s) : FID1A.CH  
 Acq On : 21 May 2020 18:10  
 Operator : DD\AJ  
 Sample : L2718-01  
 Misc :  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 FID\_C  
 ClientSampleId :  
 72-12017

Integration File: autoint1.e  
 Quant Time: May 22 01:59:14 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Alaphatic EPH 050820.M  
 Quant Title : GC Extractables  
 QLast Update : Fri May 08 16:49:27 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response | Conc Units   |
|--------------------------------|--------|----------|--------------|
| -----                          |        |          |              |
| System Monitoring Compounds    |        |          |              |
| 9) S ortho-Terphenyl (SURR)    | 11.738 | 5002468  | 41.847 ug/ml |
| Spiked Amount                  | 50.000 | Recovery | = 83.69%     |
| 12) S 1-chlorooctadecane (S... | 13.152 | 4074989  | 41.547 ug/ml |
| Spiked Amount                  | 50.000 | Recovery | = 83.09%     |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC052120AL\  
Data File : FC047544.D  
Signal(s) : FID1A.CH  
Acq On : 21 May 2020 18:10  
Operator : DD\AJ  
Sample : L2718-01  
Misc :  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
FID\_C  
ClientSampleId :  
72-12017

Integration File: autoint1.e  
Quant Time: May 22 01:59:14 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Alaphatic EPH 050820.M  
Quant Title : GC Extractables  
QLast Update : Fri May 08 16:49:27 2020  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.738 min  
Delta R.T.: -0.010 min  
Response: 5002468  
Conc: 41.85 ug/ml

Instrument :  
FID\_C  
ClientSampleId :  
72-12017



#12 1-chlorooctadecane (SURR)

R.T.: 13.152 min  
Delta R.T.: -0.009 min  
Response: 4074989  
Conc: 41.55 ug/ml