

rteres

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/27/2025
Supervised By :mohammad ahmed 05/28/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC05242
Data File : FC069003.D
Signal(s) : FID1A.ch
Acq On : 23 May 2025 17:20
Sample : 20 PPM ALIPHATIC HC STD ICV
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 052425.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.441	3.394	3.512	BB	241687	2019084	86.61%	4.800%
2	4.515	4.462	4.569	BB	226581	2011435	86.28%	4.782%
3	6.112	6.062	6.179	BB	237401	2198670	94.31%	5.227%
4	6.540	6.477	6.587	BB	201518	1996427	85.64%	4.746%
5	7.170	7.102	7.240	BB	218895	2159688	92.64%	5.135%
6	8.337	8.285	8.387	BB	189871	1941680	83.29%	4.616%
7	9.940	9.882	9.990	BB	177137	1918199	82.28%	4.560%
8	11.383	11.330	11.424	BB	169481	1894928	81.28%	4.505%
9	11.707	11.649	11.749	BB	213234	2331287	100.00%	5.543%
10	12.694	12.634	12.742	BB	158736	1851134	79.40%	4.401%
11	13.140	13.082	13.187	BB	143207	1699832	72.91%	4.041%
12	13.306	13.244	13.344	BB	153476	1825076	78.29%	4.339%
13	13.893	13.862	13.939	VB	153593	1834170	78.68%	4.361%
14	14.997	14.909	15.052	BB	143782	1874929	80.42%	4.458%
15	16.019	15.954	16.075	BB	148715	1861946	79.87%	4.427%
16	16.970	16.814	17.005	BB	144353	1857931	79.70%	4.417%
17	17.859	17.800	17.910	BB	145505	1898256	81.43%	4.513%
18	18.692	18.615	18.755	BB	141686	1919820	82.35%	4.564%
19	19.477	19.420	19.547	BB	128549	1828587	78.44%	4.347%
20	20.218	20.160	20.259	BB	123743	1783767	76.51%	4.241%
21	20.991	20.937	21.060	BB	88935	1708457	73.28%	4.062%
22	21.965	21.899	22.057	BB	63426	1646684	70.63%	3.915%
Sum of corrected areas:						42061986		

Aliphatic EPH 052425.M Tue May 27 02:03:47 2025