

rters

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC05252
Data File : FC059912.D
Signal(s) : FID1A.CH
Acq On : 25 May 2022 14:08
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Reviewed By :Yogesh Patel 05/26/2022
Supervised By :Ankita Jodhani 05/26/2022

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 052522.M
Title : GC Extractables

Signal : FID1A.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.622	2.595	2.728	BB	268581	2659429	82.69%	4.476%
2	3.620	3.598	3.741	BB	247551	2665937	82.89%	4.487%
3	5.157	5.131	5.321	BB	219668	3000496	93.29%	5.050%
4	5.574	5.538	5.685	BB	248429	2760747	85.84%	4.647%
5	6.181	6.151	6.355	BB	211732	3070799	95.48%	5.169%
6	7.323	7.285	7.431	BB	249859	2814056	87.49%	4.737%
7	8.886	8.838	8.998	BB	256522	2891738	89.91%	4.867%
8	10.293	10.255	10.361	BB	246810	2951243	91.76%	4.967%
9	10.638	10.591	10.725	BB	281805	3216303	100.00%	5.414%
10	11.572	11.515	11.638	BB	249544	2954360	91.86%	4.973%
11	12.007	11.955	12.081	BB	201380	2580655	80.24%	4.344%
12	12.170	12.128	12.255	BB	245884	2974345	92.48%	5.006%
13	12.742	12.558	12.818	BB	232551	2991888	93.02%	5.036%
14	13.819	13.725	13.905	BB	226356	2969050	92.31%	4.997%
15	14.815	14.741	14.918	BB	214490	2918410	90.74%	4.912%
16	15.742	15.605	15.851	BB	198056	2835944	88.17%	4.773%
17	16.609	16.545	16.655	BB	188855	2662602	82.78%	4.482%
18	17.421	17.358	17.498	BB	156404	2480226	77.11%	4.175%
19	18.186	18.151	18.231	BB	144873	2131512	66.27%	3.588%
20	18.904	18.867	18.941	M	121211	1976834	61.46%	3.327%
21	19.590	19.551	19.640	BBA	122247	1873834	58.26%	3.154%
22	20.239	20.201	20.301	BV	120979	2030794	63.14%	3.418%
Sum of corrected areas:						59411202		

Aliphatic EPH 052522.M Thu May 26 10:45:55 2022