

nteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2022
Supervised By :Ankita Jodhani 06/07/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC06062
Data File : FC060109.D
Signal(s) : FID1A.CH
Acq On : 07 Jun 2022 00:22
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 052722.M
Title : GC Extractables

Signal : FID1A.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.626	2.598	2.752	BB	268689	2574064	82.15%	4.325%
2	3.625	3.602	3.735	BB	253604	2613609	83.41%	4.392%
3	5.163	5.135	5.355	BB	206798	2970354	94.79%	4.991%
4	5.578	5.542	5.688	BB	248904	2748634	87.72%	4.619%
5	6.189	6.162	6.392	BB	205874	3035249	96.86%	5.100%
6	7.330	7.292	7.425	BB	254434	2820868	90.02%	4.740%
7	8.893	8.838	8.998	BB	254419	2902145	92.62%	4.877%
8	10.301	10.262	10.412	BB	243480	2956505	94.35%	4.968%
9	10.647	10.608	10.765	BB	267600	3133528	100.00%	5.265%
10	11.581	11.542	11.650	BBA	234452	2923549	93.30%	4.913%
11	12.016	11.968	12.105	BB	212974	2521243	80.46%	4.237%
12	12.179	12.135	12.278	BB	224615	2915245	93.03%	4.899%
13	12.751	12.683	12.800	M	237673	2978859	95.06%	5.006%
14	13.828	13.782	13.896	M	218550	2941783	93.88%	4.943%
15	14.827	14.778	14.873	M	224334	2861009	91.30%	4.808%
16	15.755	15.542	15.838	BB	204753	2839027	90.60%	4.771%
17	16.622	16.575	16.688	BB	191998	2776308	88.60%	4.665%
18	17.434	17.368	17.495	BB	171318	2579807	82.33%	4.335%
19	18.199	18.152	18.282	BB	151709	2298345	73.35%	3.862%
20	18.920	18.878	18.957	M	140019	2085609	66.56%	3.505%
21	19.604	19.557	19.643	M	127366	1999187	63.80%	3.359%
22	20.251	20.202	20.325	BV	126110	2035876	64.97%	3.421%
Sum of corrected areas:							59510802	

Aliphatic EPH 052722.M Tue Jun 07 11:15:32 2022