

rters

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2022
Supervised By :Ankita Jodhani 06/07/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC06062
Data File : FC060118.D
Signal(s) : FID1A.CH
Acq On : 07 Jun 2022 6:43
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 052722.M
Title : GC Extractables

Signal : FID1A.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.624	2.601	2.748	BB	258261	2584913	82.54%	4.335%
2	3.624	3.595	3.755	BB	244531	2634410	84.12%	4.418%
3	5.163	5.128	5.358	BB	212582	2991557	95.53%	5.016%
4	5.577	5.545	5.688	BB	251567	2758833	88.10%	4.626%
5	6.187	6.161	6.388	BB	212213	3060126	97.72%	5.131%
6	7.329	7.288	7.415	BB	260888	2819051	90.02%	4.727%
7	8.893	8.855	8.975	BB	256054	2896356	92.49%	4.857%
8	10.302	10.261	10.385	BB	255720	2951924	94.26%	4.950%
9	10.648	10.601	10.738	BB	263336	3131601	100.00%	5.251%
10	11.581	11.541	11.650	BBA	237036	2922077	93.31%	4.900%
11	12.016	11.961	12.098	BB	201331	2514153	80.28%	4.216%
12	12.179	12.138	12.268	BB	232334	2905537	92.78%	4.872%
13	12.751	12.702	12.808	M	233266	2971796	94.90%	4.983%
14	13.831	13.780	13.873	M	224116	2935140	93.73%	4.922%
15	14.828	14.791	14.880	BBA	207040	2837623	90.61%	4.758%
16	15.756	15.698	15.838	BB	200458	2818974	90.02%	4.727%
17	16.623	16.548	16.691	PB	185868	2798650	89.37%	4.693%
18	17.436	17.313	17.488	VB	179628	2582493	82.47%	4.331%
19	18.201	18.151	18.271	BV	152293	2300564	73.46%	3.858%
20	18.922	18.878	18.959	M	136565	2103847	67.18%	3.528%
21	19.606	19.564	19.652	M	128549	1999880	63.86%	3.354%
22	20.253	20.201	20.335	BV	123107	2115421	67.55%	3.547%
Sum of corrected areas:							59634925	

Aliphatic EPH 052722.M Tue Jun 07 11:17:01 2022