

rters

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/09/2025
Supervised By :mohammad ahmed 06/10/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC06062
Data File : FC069110.D
Signal(s) : FID1A.ch
Acq On : 06 Jun 2025 11:58
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 052425.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.403	3.370	3.472	BB	265731	2218726	89.18%	4.957%
2	4.478	4.420	4.532	BB	243483	2220027	89.23%	4.960%
3	6.072	6.015	6.134	BB	265785	2413029	96.99%	5.392%
4	6.503	6.440	6.550	BB	218317	2184475	87.80%	4.881%
5	7.132	7.064	7.207	BB	241211	2360558	94.88%	5.274%
6	8.303	8.244	8.365	BB	202270	2091121	84.05%	4.672%
7	9.908	9.845	9.955	BB	182139	2022987	81.31%	4.520%
8	11.354	11.330	11.399	BB	173065	1924848	77.37%	4.301%
9	11.675	11.617	11.734	BB	231609	2487917	100.00%	5.559%
10	12.666	12.627	12.717	BB	162627	1917765	77.08%	4.285%
11	13.112	13.055	13.165	BB	146610	1793537	72.09%	4.007%
12	13.279	13.232	13.334	BB	157103	1899006	76.33%	4.243%
13	13.867	13.790	13.919	BB	147507	1891531	76.03%	4.226%
14	14.972	14.934	15.009	BB	144266	1874375	75.34%	4.188%
15	15.996	15.942	16.050	BB	147270	1874217	75.33%	4.188%
16	16.949	16.902	16.989	BB	144231	1906903	76.65%	4.261%
17	17.838	17.800	17.875	BB	146977	2035050	81.80%	4.547%
18	18.672	18.609	18.732	BB	146382	2093368	84.14%	4.677%
19	19.459	19.420	19.497	BB	138464	1994780	80.18%	4.457%
20	20.201	20.160	20.240	BV	133255	1944706	78.17%	4.345%
21	20.969	20.892	21.040	BB	100788	1857015	74.64%	4.149%
22	21.937	21.890	22.022	BB	73258	1749499	70.32%	3.909%
Sum of corrected areas:						44755439		

Aliphatic EPH 052425.M Sat Jun 07 03:38:45 2025