

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD

rteres

Area Percent Report

Manual Integrations
APPROVED

Ankita
6/10/2021 3:16:16 PM

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC060921AL\
Data File : FC054172.D
Signal(s) : FID1A.CH
Acq On : 10 Jun 2021 7:31
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 052721.M
Title : GC Extractables

Signal : FID1A.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.162	3.135	3.258	BB	313052	2595459	87.29%	4.598%
2	4.218	4.188	4.298	BB	298058	2658320	89.41%	4.709%
3	5.785	5.732	5.925	BB	290302	2921441	98.26%	5.176%
4	6.227	6.158	6.305	BB	278244	2717133	91.39%	4.814%
5	6.838	6.782	6.988	BB	293539	2973254	100.00%	5.267%
6	8.015	7.935	8.102	BB	261893	2726473	91.70%	4.830%
7	9.610	9.538	9.685	BB	251823	2667673	89.72%	4.726%
8	11.049	10.975	11.115	BB	238921	2709235	91.12%	4.800%
9	11.353	11.278	11.442	BB	262550	2791700	93.89%	4.946%
10	12.356	12.275	12.425	BB	221987	2655116	89.30%	4.704%
11	12.797	12.645	12.875	BB	182507	2164707	72.81%	3.835%
12	12.967	12.895	13.038	BB	221130	2587078	87.01%	4.583%
13	13.552	13.519	13.608	VB	211050	2560777	86.13%	4.537%
14	14.654	14.562	14.722	BB	192429	2515574	84.61%	4.456%
15	15.674	15.588	15.738	BB	183960	2427473	81.64%	4.300%
16	16.623	16.445	16.685	BB	182544	2458139	82.68%	4.355%
17	17.509	17.402	17.549	BB	187646	2541869	85.49%	4.503%
18	18.342	18.288	18.392	BB	179493	2565822	86.30%	4.546%
19	19.125	19.065	19.195	BB	169475	2476770	83.30%	4.388%
20	19.864	19.822	19.915	BB	160910	2324593	78.18%	4.118%
21	20.572	20.515	20.617	M	134434	2189581	73.64%	3.879%
22	21.402	21.295	21.492	BB	98571	2219180	74.64%	3.931%
Sum of corrected areas:						56447367		

Alaphatic EPH 052721.M Thu Jun 10 13:18:50 2021