

rteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent

Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data
Data File : FC066598.D
Signal(s) : FID1A.ch
Acq On : 13 Jun 2024 02:20
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Reviewed By :Yogesh Patel 06/13/2024
Supervised By :Ankita Jodhani 06/13/2024

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH
052824.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.303	3.250	3.345	BB	318467	2707403	88.89%	4.584%
2	4.372	4.318	4.428	BB	308441	2792856	91.70%	4.729%
3	5.961	5.902	6.030	BB	334473	3045673	100.00%	5.157%
4	6.395	6.340	6.445	BB	278428	2764131	90.76%	4.680%
5	7.019	6.970	7.085	BB	309038	2923268	95.98%	4.950%
6	8.191	8.130	8.248	BB	255298	2699043	88.62%	4.570%
7	9.794	9.747	9.845	BB	242685	2753354	90.40%	4.662%
8	11.237	11.200	11.280	BB	235690	2739809	89.96%	4.639%
9	11.554	11.520	11.598	BB	265682	2868231	94.17%	4.857%
10	12.548	12.507	12.588	BB	212981	2602412	85.45%	4.406%
11	12.992	12.952	13.053	BB	187043	2197760	72.16%	3.721%
12	13.160	13.120	13.197	BV	208737	2553279	83.83%	4.323%
13	13.747	13.623	13.790	BB	198864	2550663	83.75%	4.319%
14	14.850	14.760	14.900	BB	192842	2536221	83.27%	4.294%
15	15.871	15.813	15.917	BB	184736	2504053	82.22%	4.240%
16	16.822	16.398	16.860	BV	190759	2513840	82.54%	4.256%
17	17.711	17.622	17.748	PV	196674	2767900	90.88%	4.687%
18	18.544	18.450	18.580	BV	183005	2745461	90.14%	4.649%
19	19.328	19.250	19.363	BV	188222	2659983	87.34%	4.504%
20	20.068	20.015	20.101	VV	178355	3016700	99.05%	5.108%
21	20.805	20.764	20.843	PV	126926	2332023	76.57%	3.949%
22	21.717	21.660	21.800	PBA	100946	2785444	91.46%	4.716%
Sum of corrected areas:						59059508		

Aliphatic EPH 052824.M Thu Jun 13 04:40:45 2024

Page 1