

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061323AL\
 Data File : FC063831.D
 Signal(s) : FID1A.ch
 Acq On : 13 Jun 2023 16:10
 Operator : YP/AJ
 Sample : PB153398BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jun 14 01:00:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 052723.M
 Quant Title : GC Extractables
 QLast Update : Sat May 27 06:12:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.540	8424977	46.168 ug/ml
Spiked Amount	50.000	Recovery	= 92.34%
12) S 1-chlorooctadecane (S...	12.974	6141089	48.063 ug/ml
Spiked Amount	50.000	Recovery	= 96.13%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061323AL\
Data File : FC063831.D
Signal(s) : FID1A.ch
Acq On : 13 Jun 2023 16:10
Operator : YP/AJ
Sample : PB153398BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jun 14 01:00:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 052723.M
Quant Title : GC Extractables
QLast Update : Sat May 27 06:12:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.540 min
Delta R.T.: 0.007 min
Response: 8424977
Conc: 46.17 ug/ml

Instrument :
FID_C
ClientSampleId :



#12 1-chlorooctadecane (SURR)

R.T.: 12.974 min
Delta R.T.: 0.007 min
Response: 6141089
Conc: 48.06 ug/ml