

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061422AL\
 Data File : FC060249.D
 Signal(s) : FID1A.CH
 Acq On : 15 Jun 2022 6:17
 Operator : UA/NP
 Sample : N3336-01
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jun 15 07:49:14 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 052722.M
 Quant Title : GC Extractables
 QLast Update : Sat May 28 08:22:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	10.650	4774246	30.678 ug/ml
Spiked Amount 50.000		Recovery =	61.36%
12) S 1-chlorooctadecane (S...	12.024	6213449	50.113 ug/ml
Spiked Amount 50.000		Recovery =	100.23%
Target Compounds			
14) T n-Tetracosane (C24)	13.837	471950	3.363 ug/ml
16) T n-Octacosane (C28)	15.716f	400424	3.075 ug/ml
19) T n-Tetratriacontane (C34)	18.261	109542	1.074 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061422AL\
Data File : FC060249.D
Signal(s) : FID1A.CH
Acq On : 15 Jun 2022 6:17
Operator : UA/NP
Sample : N3336-01
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jun 15 07:49:14 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 052722.M
Quant Title : GC Extractables
QLast Update : Sat May 28 08:22:04 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 10.650 min
Delta R.T.: -0.006 min
Response: 4774246
Conc: 30.68 ug/ml

Instrument :
FID_C
ClientSampleId :




#12 1-chlorooctadecane (SURR)

R.T.: 12.024 min
Delta R.T.: 0.000 min
Response: 6213449
Conc: 50.11 ug/ml




#14 n-Tetracosane (C24)

R.T.: 13.837 min
Delta R.T.: -0.002 min
Response: 471950
Conc: 3.36 ug/ml



#16 n-Octacosane (C28)

R.T.: 15.716 min
Delta R.T.: -0.050 min
Response: 400424
Conc: 3.07 ug/ml



#19 n-Tetratriacontane (C34)

R.T.: 18.261 min
Delta R.T.: 0.048 min
Response: 109542
Conc: 1.07 ug/ml

Instrument :
FID_C
ClientSampleId :