

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061920AL\
Data File : FC047951.D
Signal(s) : FID1A.CH
Acq On : 19 Jun 2020 14:48
Operator : DD\AJ
Sample : MECL2
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
MECL2

Integration File: autoint1.e
Quant Time: Jun 20 01:19:21 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 060920.M
Quant Title : GC Extractables
QLast Update : Wed Jun 10 08:39:57 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
Target Compounds				
19) T	n-Tetratriacontane (C34)	19.435	155495	1.698 ug/ml
21) T	n-Octatriacontane (C38)	20.912	500792	6.289 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061920AL\
Data File : FC047951.D
Signal(s) : FID1A.CH
Acq On : 19 Jun 2020 14:48
Operator : DD\AJ
Sample : MECL2
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
MECL2

Integration File: autoint1.e
Quant Time: Jun 20 01:19:21 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 060920.M
Quant Title : GC Extractables
QLast Update : Wed Jun 10 08:39:57 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 0.000 min
Exp R.T. : 11.739 min
Response: 0
Conc: N.D.

Instrument :
FID_C
ClientSampleId :
MECL2




#12 1-chlorooctadecane (SURR)

R.T.: 0.000 min
Exp R.T. : 13.150 min
Response: 0
Conc: N.D.



#19 n-Tetratriacontane (C34)

R.T.: 19.435 min
Delta R.T.: 0.000 min
Response: 155495
Conc: 1.70 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 20.912 min
Delta R.T.: -0.011 min
Response: 500792
Conc: 6.29 ug/ml