

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC062321AL\
 Data File : FC054390.D
 Signal(s) : FID1A.CH
 Acq On : 23 Jun 2021 11:10
 Operator : DD\AJ
 Sample : M2778-05DL 5X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jun 24 02:18:06 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 061021.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 11 04:43:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.340	1274304	8.391 ug/ml
Spiked Amount 50.000		Recovery =	16.78%
12) S 1-chlorooctadecane (S...	12.784	911502	7.738 ug/ml
Spiked Amount 50.000		Recovery =	15.48%
Target Compounds			
7) T n-Hexadecane (C16)	9.550f	184394	1.363 ug/ml
10) T n-Eicosane (C20)	12.391	198381	1.429 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC062321AL\
Data File : FC054390.D
Signal(s) : FID1A.CH
Acq On : 23 Jun 2021 11:10
Operator : DD\AJ
Sample : M2778-05DL 5X
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jun 24 02:18:06 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061021.M
Quant Title : GC Extractables
QLast Update : Fri Jun 11 04:43:57 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#7 n-Hexadecane (C16)


R.T.: 9.550 min
Delta R.T.: -0.058 min
Response: 184394
Conc: 1.36 ug/ml

Instrument :
FID_C
ClientSampleId :




#9 ortho-Terphenyl (SURR)

R.T.: 11.340 min
Delta R.T.: -0.010 min
Response: 1274304
Conc: 8.39 ug/ml



#10 n-Eicosane (C20)

R.T.: 12.391 min
Delta R.T.: 0.038 min
Response: 198381
Conc: 1.43 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 12.784 min
Delta R.T.: -0.011 min
Response: 911502
Conc: 7.74 ug/ml