

rters

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/24/2022
Supervised By :Ankita Jodhani 06/24/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC06232
Data File : FC060411.D
Signal(s) : FID1A.CH
Acq On : 23 Jun 2022 10:21
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 052722.M
Title : GC Extractables

Signal : FID1A.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.619	2.591	2.681	BB	273496	2534900	90.45%	4.629%
2	3.621	3.595	3.681	BB	254344	2564975	91.52%	4.684%
3	5.165	5.135	5.238	BB	209368	2611075	93.17%	4.768%
4	5.577	5.548	5.661	BB	251046	2719766	97.05%	4.967%
5	6.190	6.161	6.288	BB	207736	2748659	98.08%	5.020%
6	7.331	7.288	7.405	BB	251600	2725010	97.23%	4.977%
7	8.896	8.861	8.971	BB	233749	2733016	97.52%	4.991%
8	10.308	10.265	10.378	BB	234174	2689946	95.98%	4.912%
9	10.654	10.611	10.728	BB	247219	2802565	100.00%	5.118%
10	11.588	11.548	11.655	BB	225958	2587664	92.33%	4.726%
11	12.024	11.981	12.095	BB	185397	2197296	78.40%	4.013%
12	12.187	12.148	12.261	BB	199571	2546260	90.85%	4.650%
13	12.761	12.715	12.821	BB	205651	2503232	89.32%	4.571%
14	13.841	13.778	13.905	BB	183071	2433108	86.82%	4.443%
15	14.840	14.778	14.895	BB	177331	2369474	84.55%	4.327%
16	15.770	15.705	15.821	BB	169172	2398274	85.57%	4.380%
17	16.638	16.565	16.675	BB	172270	2472447	88.22%	4.515%
18	17.452	17.398	17.511	BB	166421	2471273	88.18%	4.513%
19	18.220	18.171	18.258	BB	148273	2300215	82.08%	4.201%
20	18.941	18.888	18.978	BB	139638	2184962	77.96%	3.990%
21	19.627	19.551	19.675	BB	142563	2145533	76.56%	3.918%
22	20.277	20.242	20.313	M	122029	2017888	72.00%	3.685%
Sum of corrected areas:						54757536		

Aliphatic EPH 052722.M Fri Jun 24 11:52:32 2022