

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC062521AL\
 Data File : FC054456.D
 Signal(s) : FID1A.CH
 Acq On : 26 Jun 2021 3:03
 Operator : DD\AJ
 Sample : M2819-02DL 5X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jun 26 06:29:57 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 062521.M
 Quant Title : GC Extractables
 QLast Update : Sat Jun 26 02:10:35 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.329	1098620	7.760 ug/ml
Spiked Amount 50.000		Recovery =	15.52%
12) S 1-chlorooctadecane (S...	12.773	1083866	9.723 ug/ml
Spiked Amount 50.000		Recovery =	19.45%
Target Compounds			
10) T n-Eicosane (C20)	12.378	1684677	12.947 ug/ml
14) T n-Tetracosane (C24)	14.594	702330	5.188 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC062521AL\
Data File : FC054456.D
Signal(s) : FID1A.CH
Acq On : 26 Jun 2021 3:03
Operator : DD\AJ
Sample : M2819-02DL 5X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jun 26 06:29:57 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 062521.M
Quant Title : GC Extractables
QLast Update : Sat Jun 26 02:10:35 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 11.329 min
Delta R.T.: 0.000 min
Response: 1098620
Conc: 7.76 ug/ml

Instrument :
FID_C
ClientSampleId :




#10 n-Eicosane (C20)

R.T.: 12.378 min
Delta R.T.: 0.046 min
Response: 1684677
Conc: 12.95 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 12.773 min
Delta R.T.: 0.000 min
Response: 1083866
Conc: 9.72 ug/ml



#14 n-Tetracosane (C24)

R.T.: 14.594 min
Delta R.T.: -0.034 min
Response: 702330
Conc: 5.19 ug/ml