

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC062623AL\
 Data File : FC063989.D
 Signal(s) : FID1A.ch
 Acq On : 26 Jun 2023 09:03
 Operator : YP/AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Jun 27 03:54:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 062423.M
 Quant Title : GC Extractables
 QLast Update : Sun Jun 25 04:03:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.523	3715575	19.136 ug/ml
Spiked Amount 50.000		Recovery =	38.27%
12) S 1-chlorooctadecane (S...	12.959	2647685	19.030 ug/ml
Spiked Amount 50.000		Recovery =	38.06%
Target Compounds			
1) T n-Nonane (C9)	3.304	3432236	19.244 ug/ml
2) T n-Decane (C10)	4.367	3478886	19.225 ug/ml
3) T A~Naphthalene (C11.7)	5.948	3582078	19.204 ug/ml
4) T n-Dodecane (C12)	6.382	3460439	19.209 ug/ml
5) T A~2-methylnaphthalene...	7.004	3384034	19.229 ug/ml
6) T n-Tetradecane (C14)	8.174	3314545	19.160 ug/ml
7) T n-Hexadecane (C16)	9.772	3317550	19.098 ug/ml
8) T n-Octadecane (C18)	11.210	3295795	19.123 ug/ml
10) T n-Eicosane (C20)	12.516	3100281	18.991 ug/ml
11) T n-Heneicosane (C21)	13.126	3026070	18.887 ug/ml
13) T n-Docosane (C22)	13.711	2976223	18.940 ug/ml
14) T n-Tetracosane (C24)	14.811	2918272	18.514 ug/ml
15) T n-Hexacosane (C26)	15.828	2853197	18.885 ug/ml
16) T n-Octacosane (C28)	16.775	2801742	18.594 ug/ml
17) T n-Tricontane (C30)	17.660	2885455	18.481 ug/ml
18) T n-Dotriacontane (C32)	18.490	2884677	18.260 ug/ml
19) T n-Tetratriacontane (C34)	19.270	2858459	18.576 ug/ml
20) T n-Hexatriacontane (C36)	20.007	2756774	18.673 ug/ml
21) T n-Octatriacontane (C38)	20.732	2800248	18.904 ug/ml
22) T n-Tetracontane (C40)	21.622	2754261	18.896 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC062623AL\
Data File : FC063989.D
Signal(s) : FID1A.ch
Acq On : 26 Jun 2023 09:03
Operator : YP/AJ
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
Quant Time: Jun 27 03:54:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 062423.M
Quant Title : GC Extractables
QLast Update : Sun Jun 25 04:03:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

