

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC062623AL\
 Data File : FC063997.D
 Signal(s) : FID1A.ch
 Acq On : 26 Jun 2023 19:25
 Operator : YP/AJ
 Sample : 03412-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 DUCT-1

Integration File: autoint1.e
 Quant Time: Jun 27 03:55:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 062423.M
 Quant Title : GC Extractables
 QLast Update : Sun Jun 25 04:03:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.962	5301362	38.104 ug/ml
Spiked Amount	50.000	Recovery	= 76.21%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC062623AL\
Data File : FC063997.D
Signal(s) : FID1A.ch
Acq On : 26 Jun 2023 19:25
Operator : YP/AJ
Sample : 03412-01
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
DUCT-1

Integration File: autoint1.e
Quant Time: Jun 27 03:55:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 062423.M
Quant Title : GC Extractables
QLast Update : Sun Jun 25 04:03:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min
Exp R.T.: 11.525 min
Response: 0
Conc: N.D.

Instrument :
FID_C
ClientSampleId :
DUCT-1



#12 1-chlorooctadecane (SURR)

R.T.: 12.962 min
Delta R.T.: 0.002 min
Response: 5301362
Conc: 38.10 ug/ml