

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC062723AL\
 Data File : FC064024.D
 Signal(s) : FID1A.ch
 Acq On : 28 Jun 2023 01:00
 Operator : YP/AJ
 Sample : 03415-07
 Misc :
 ALS Vial : 90 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 ESI-TP1G

Integration File: autoint1.e
 Quant Time: Jun 28 02:53:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 062423.M
 Quant Title : GC Extractables
 QLast Update : Sun Jun 25 04:03:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.526	6546626	33.717 ug/ml
Spiked Amount	50.000	Recovery	= 67.43%
12) S 1-chlorooctadecane (S...	12.961	5064803	36.403 ug/ml
Spiked Amount	50.000	Recovery	= 72.81%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC062723AL\
Data File : FC064024.D
Signal(s) : FID1A.ch
Acq On : 28 Jun 2023 01:00
Operator : YP/AJ
Sample : 03415-07
Misc :
ALS Vial : 90 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
ESI-TP1G

Integration File: autoint1.e
Quant Time: Jun 28 02:53:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 062423.M
Quant Title : GC Extractables
QLast Update : Sun Jun 25 04:03:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.526 min
Delta R.T.: 0.001 min
Response: 6546626
Conc: 33.72 ug/ml

Instrument :
FID_C
ClientSampleId :
ESI-TP1G



#12 1-chlorooctadecane (SURR)

R.T.: 12.961 min
Delta R.T.: 0.000 min
Response: 5064803
Conc: 36.40 ug/ml