

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC062921AL\
 Data File : FC054490.D
 Signal(s) : FID1A.CH
 Acq On : 29 Jun 2021 11:54
 Operator : DD\AJ
 Sample : M2812-07
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 KG-WATER

Integration File: sample.E
 Quant Time: Jun 29 13:31:27 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 062521.M
 Quant Title : GC Extractables
 QLast Update : Sat Jun 26 02:10:35 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.781	3232335	28.996 ug/ml
Spiked Amount 50.000		Recovery =	57.99%
Target Compounds			

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC062921AL\
Data File : FC054490.D
Signal(s) : FID1A.CH
Acq On : 29 Jun 2021 11:54
Operator : DD\AJ
Sample : M2812-07
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
KG-WATER

Integration File: sample.E
Quant Time: Jun 29 13:31:27 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 062521.M
Quant Title : GC Extractables
QLast Update : Sat Jun 26 02:10:35 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min
Exp R.T.: 11.329 min
Response: 0
Conc: N.D.

Instrument :
FID_C
ClientSampleId :
KG-WATER



#12 1-chlorooctadecane (SURR)

R.T.: 12.781 min
Delta R.T.: 0.008 min
Response: 3232335
Conc: 29.00 ug/ml