

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 07/05/2023
Supervised By :Ankita Jodhani 07/05/2023

rteres

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\F0
Data File : FC064046.D
Signal(s) : FID1A.ch
Acq On : 04 Jul 2023 00:25
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 062423.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.306	3.282	3.356	BB	423947	3587897	89.44%	4.865%
2	4.368	4.338	4.424	BB	398945	3680059	91.74%	4.990%
3	5.949	5.917	6.037	BB	387251	3801223	94.76%	5.154%
4	6.381	6.352	6.438	BB	377441	3671389	91.52%	4.978%
5	7.004	6.961	7.102	BB	337269	3604434	89.85%	4.888%
6	8.173	8.139	8.238	BB	331406	3557064	88.67%	4.823%
7	9.770	9.735	9.837	BB	319520	3583910	89.34%	4.860%
8	11.209	11.175	11.284	BB	315244	3583143	89.32%	4.859%
9	11.523	11.487	11.572	BB	378631	4011404	100.00%	5.439%
10	12.514	12.477	12.566	BB	282455	3378400	84.22%	4.581%
11	12.958	12.922	13.011	BB	236707	2898887	72.27%	3.931%
12	13.125	13.089	13.177	BB	278222	3300554	82.28%	4.476%
13	13.709	13.672	13.764	BB	258709	3249211	81.00%	4.406%
14	14.808	14.766	14.864	BB	250912	3190065	79.52%	4.326%
15	15.826	15.778	15.878	BB	239454	3115558	77.67%	4.225%
16	16.772	16.723	16.830	BB	219135	3071430	76.57%	4.165%
17	17.658	17.599	17.712	BB	223572	3190173	79.53%	4.326%
18	18.486	18.433	18.540	BB	228028	3194468	79.63%	4.332%
19	19.267	19.211	19.322	BB	219082	3137752	78.22%	4.255%
20	20.006	19.975	20.047	BB	215183	2953943	73.64%	4.006%
21	20.729	20.664	20.777	BB	179526	3016569	75.20%	4.090%
22	21.615	21.539	21.700	BB	125271	2969545	74.03%	4.027%
Sum of corrected areas:						73747075		

Aliphatic EPH 062423.M Wed Jul 05 01:42:29 2023