

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 07/07/2023
Supervised By :Ankita Jodhani 07/07/2023

rteres

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\F0
Data File : FC064064.D
Signal(s) : FID1A.ch
Acq On : 07 Jul 2023 01:49
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 062423.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.307	3.267	3.355	BB	465041	3932580	93.02%	5.060%
2	4.369	4.324	4.423	BB	434600	3986682	94.30%	5.129%
3	5.949	5.907	6.037	BB	424032	4145892	98.06%	5.334%
4	6.382	6.346	6.433	BB	402165	3961274	93.70%	5.097%
5	7.004	6.972	7.105	BB	381262	3944393	93.30%	5.075%
6	8.173	8.139	8.237	BB	356050	3813929	90.21%	4.907%
7	9.770	9.720	9.828	BB	341436	3807909	90.07%	4.899%
8	11.209	11.172	11.272	BB	324012	3763419	89.02%	4.842%
9	11.522	11.487	11.573	BB	398115	4227738	100.00%	5.440%
10	12.515	12.478	12.567	BB	294871	3539981	83.73%	4.555%
11	12.958	12.922	13.007	BB	259830	3047854	72.09%	3.921%
12	13.124	13.088	13.172	BB	291358	3442183	81.42%	4.429%
13	13.709	13.671	13.761	BB	280213	3393912	80.28%	4.367%
14	14.808	14.771	14.861	BB	260611	3338992	78.98%	4.296%
15	15.824	15.785	15.875	BB	245632	3275851	77.48%	4.215%
16	16.772	16.722	16.822	BB	239140	3202239	75.74%	4.120%
17	17.657	17.604	17.703	BB	234073	3273236	77.42%	4.211%
18	18.486	18.433	18.532	BB	233565	3248673	76.84%	4.180%
19	19.267	19.209	19.310	BB	217763	3208670	75.90%	4.128%
20	20.004	19.958	20.045	M	216526	3093891	73.18%	3.981%
21	20.727	20.646	20.791	BB	180155	3084205	72.95%	3.968%
22	21.612	21.535	21.724	BB	127706	2989038	70.70%	3.846%
Sum of corrected areas:						77722540		

Aliphatic EPH 062423.M Fri Jul 07 03:34:23 2023