

rteres

Instrument :

FID_C

LabSampleId :

20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 07/08/2022

Supervised By :Ankita Jodhani 07/08/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC07072
Data File : FC060630.D
Signal(s) : FID1A.ch
Acq On : 08 Jul 2022 02:22
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 062422.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.624	2.600	2.739	BB	333597	3213351	89.30%	4.728%
2	3.619	3.594	3.732	BB	311666	3228536	89.72%	4.750%
3	5.157	5.119	5.335	BB	245125	3511341	97.58%	5.167%
4	5.568	5.539	5.677	BB	312248	3404228	94.60%	5.009%
5	6.180	6.095	6.359	BB	243250	3543305	98.47%	5.214%
6	7.318	7.282	7.412	BB	308232	3431609	95.36%	5.049%
7	8.880	8.844	8.955	BB	308979	3455481	96.03%	5.084%
8	10.288	10.217	10.360	BB	304327	3438178	95.55%	5.059%
9	10.635	10.594	10.719	BB	304725	3598456	100.00%	5.295%
10	11.567	11.515	11.627	BB	279245	3344927	92.95%	4.922%
11	12.003	11.945	12.067	BB	232904	2850641	79.22%	4.194%
12	12.165	12.097	12.220	BB	264473	3305027	91.85%	4.863%
13	12.737	12.672	12.795	BB	258434	3251827	90.37%	4.785%
14	13.815	13.749	13.869	BB	244192	3115574	86.58%	4.584%
15	14.811	14.739	14.865	BB	218701	2963196	82.35%	4.360%
16	15.740	15.672	15.784	BB	216090	2840721	78.94%	4.180%
17	16.606	16.539	16.647	BB	192546	2768343	76.93%	4.073%
18	17.418	17.340	17.462	BB	180462	2659199	73.90%	3.913%
19	18.183	18.110	18.219	BB	167331	2490962	69.22%	3.665%
20	18.902	18.850	18.960	BBA	155443	2408405	66.93%	3.544%
21	19.585	19.302	19.669	BV	155255	2705506	75.19%	3.981%
22	20.234	20.180	20.277	BB	150013	2434542	67.66%	3.582%
Sum of corrected areas:						67963353		

Aliphatic EPH 062422.M Fri Jul 08 19:10:00 2022