

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC071023AL\
 Data File : FC064074.D
 Signal(s) : FID1A.ch
 Acq On : 10 Jul 2023 10:38
 Operator : YP/AJ
 Sample : PB154061BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jul 11 02:56:52 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 062423.M
 Quant Title : GC Extractables
 QLast Update : Sun Jun 25 04:03:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.523	7650874	39.404 ug/ml
Spiked Amount 50.000		Recovery =	78.81%
12) S 1-chlorooctadecane (S...	12.959	5864936	42.154 ug/ml
Spiked Amount 50.000		Recovery =	84.31%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC071023AL\
Data File : FC064074.D
Signal(s) : FID1A.ch
Acq On : 10 Jul 2023 10:38
Operator : YP/AJ
Sample : PB154061BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jul 11 02:56:52 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 062423.M
Quant Title : GC Extractables
QLast Update : Sun Jun 25 04:03:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.523 min
Delta R.T.: -0.002 min
Response: 7650874
Conc: 39.40 ug/ml

Instrument :
FID_C
ClientSampleId :



#12 1-chlorooctadecane (SURR)

R.T.: 12.959 min
Delta R.T.: -0.002 min
Response: 5864936
Conc: 42.15 ug/ml