

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC071124AL\  
 Data File : FC066766.D  
 Signal(s) : FID1A.ch  
 Acq On : 11 Jul 2024 20:24  
 Operator : YP/AJ  
 Sample : P3207-04  
 Misc :  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 FID\_C  
 ClientSampleId :  
 RBR2520700-E2

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 07/12/2024  
 Supervised By :Ankita Jodhani 07/12/2024

Integration File: autoint1.e  
 Quant Time: Jul 12 02:53:40 2024  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 061724.M  
 Quant Title : GC Extractables  
 QLast Update : Mon Jun 17 13:44:39 2024  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units    |
|--------------------------------|--------|------------|---------------|
| -----                          |        |            |               |
| System Monitoring Compounds    |        |            |               |
| 9) S ortho-Terphenyl (SURR)    | 11.536 | 6941109    | 40.241 ug/mlm |
| Spiked Amount 50.000           |        | Recovery = | 80.48%        |
| 12) S 1-chlorooctadecane (S... | 12.974 | 5185915    | 37.585 ug/mlm |
| Spiked Amount 50.000           |        | Recovery = | 75.17%        |

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC071124AL\  
Data File : FC066766.D  
Signal(s) : FID1A.ch  
Acq On : 11 Jul 2024 20:24  
Operator : YP/AJ  
Sample : P3207-04  
Misc :  
ALS Vial : 15 Sample Multiplier: 1


Instrument :  
FID\_C  
ClientSampleId :  
RBR2520700-E2


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 07/12/2024  
Supervised By :Ankita Jodhani 07/12/2024

Integration File: autoint1.e  
Quant Time: Jul 12 02:53:40 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 061724.M  
Quant Title : GC Extractables  
QLast Update : Mon Jun 17 13:44:39 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.536 min  
Delta R.T.: -0.008 min  
Response: 6941109  
Conc: 40.24 ug/ml

Instrument :  
FID\_C  
ClientSampleId :  
RBR2520700-E2

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 07/12/2024  
Supervised By :Ankita Jodhani 07/12/2024



#12 1-chlorooctadecane (SURR)

R.T.: 12.974 min  
Delta R.T.: -0.009 min  
Response: 5185915  
Conc: 37.58 ug/ml m