

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC071818AL\
 Data File : FC035452.D
 Signal(s) : FID1A.CH
 Acq On : 18 Jul 2018 15:57
 Operator : AJ\SJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: sample.E
 Quant Time: Jul 19 00:56:14 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_C\Method\Alaphatic EPH 062518.M
 Quant Title : GC Extractables
 QLast Update : Mon Jun 25 14:27:40 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	83.857	86.867 E3	-3.6	100	-0.01
2 T	n-Decane (C10)	83.813	90.953 E3	-8.5	106	-0.01
3 T	A-Naphthalene (C11.7)	94.890	119.380 E3	-25.8#	123	-0.01
4 T	n-Dodecane (C12)	86.311	100.131 E3	-16.0	114	0.00
5 T	A-2-methylnaphthalene (C12.)	96.295	122.924 E3	-27.7#	125	0.00
6 T	n-Tetradecane (C14)	89.174	106.217 E3	-19.1	119	0.00
7 T	n-Hexadecane (C16)	91.744	105.750 E3	-15.3	117	0.00
8 T	n-Octadecane (C18)	94.376	104.520 E3	-10.7	111	0.00
9 S	ortho-Terphenyl (SURR)	102.960	122.302 E3	-18.8	121	0.00
10 T	n-Eicosane (C20)	96.430	106.229 E3	-10.2	113	0.00
11 T	n-Heneicosane (C21)	94.999	105.214 E3	-10.8	113	0.00
12 S	1-chlorooctadecane (SURR)	80.043	89.765 E3	-12.1	114	0.00
13 T	n-Docosane (C22)	95.442	106.306 E3	-11.4	114	0.00
14 T	n-Tetracosane (C24)	95.452	105.108 E3	-10.1	112	0.00
15 T	n-Hexacosane (C26)	94.536	104.494 E3	-10.5	112	0.00
16 T	n-Octacosane (C28)	94.019	103.893 E3	-10.5	111	0.00
17 T	n-Tricontane (C30)	91.885	102.554 E3	-11.6	112	0.00
18 T	n-Dotriacontane (C32)	88.793	101.268 E3	-14.0	114	0.00
19 T	n-Tetratriacontane (C34)	90.481	98.032 E3	-8.3	109	0.00
20 T	n-Hexatriacontane (C36)	89.661	89.962 E3	-0.3	100	0.00
21 T	n-Octatriacontane (C38)	92.226	101.703 E3	-10.3	112	0.00
22 T	n-Tetracontane (C40)	93.982	101.588 E3	-8.1	108	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0