

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC072822AL\  
 Data File : FC060879.D  
 Signal(s) : FID1A.ch  
 Acq On : 28 Jul 2022 16:36  
 Operator : YP/AJ  
 Sample : N3925-01  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 FID\_C  
 ClientSampleId :  
 JC-03-072722-A

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 07/29/2022  
 Supervised By :Ankita Jodhani 07/29/2022

Integration File: autoint1.e  
 Quant Time: Jul 28 22:50:50 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 072522.M  
 Quant Title : GC Extractables  
 QLast Update : Mon Jul 25 19:40:54 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.    | Response   | Conc Units     |
|--------------------------------|---------|------------|----------------|
| -----                          |         |            |                |
| System Monitoring Compounds    |         |            |                |
| 9) S ortho-Terphenyl (SURR)    | 10.636  | 6463733    | 39.387 ug/ml   |
| Spiked Amount 50.000           |         | Recovery = | 78.77%         |
| 12) S 1-chlorooctadecane (S... | 12.064f | 102954682  | 775.525 ug/mlm |
| Spiked Amount 50.000           |         | Recovery = | 1551.05%       |

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC072822AL\  
Data File : FC060879.D  
Signal(s) : FID1A.ch  
Acq On : 28 Jul 2022 16:36  
Operator : YP/AJ  
Sample : N3925-01  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
FID\_C  
ClientSampleId :  
JC-03-072722-A

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 07/29/2022  
Supervised By :Ankita Jodhani 07/29/2022

Integration File: autoint1.e  
Quant Time: Jul 28 22:50:50 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 072522.M  
Quant Title : GC Extractables  
QLast Update : Mon Jul 25 19:40:54 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um

