

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC072823AL\
 Data File : FC064282.D
 Signal(s) : FID1A.ch
 Acq On : 28 Jul 2023 14:38
 Operator : YP/AJ
 Sample : 03771-01DL 5X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jul 29 01:46:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 072323.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 24 01:54:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.951	539185	4.276 ug/ml
Spiked Amount 50.000		Recovery =	8.55%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC072823AL\
Data File : FC064282.D
Signal(s) : FID1A.ch
Acq On : 28 Jul 2023 14:38
Operator : YP/AJ
Sample : 03771-01DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jul 29 01:46:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 072323.M
Quant Title : GC Extractables
QLast Update : Mon Jul 24 01:54:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min
Exp R.T.: 11.513 min
Response: 0
Conc: N.D.

Instrument :
FID_C
ClientSampleId :



#12 1-chlorooctadecane (SURR)

R.T.: 12.951 min
Delta R.T.: 0.003 min
Response: 539185
Conc: 4.28 ug/ml