

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC072825AL\  
 Data File : FC069547.D  
 Signal(s) : FID1A.ch  
 Acq On : 28 Jul 2025 15:18  
 Operator : YP/AJ  
 Sample : PB169021BL  
 Misc :  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 FID\_C  
 ClientSampleId :  
 PB169021BL

Integration File: autoint1.e  
 Quant Time: Jul 29 01:15:54 2025  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 071525.M  
 Quant Title : GC Extractables  
 QLast Update : Tue Jul 15 12:35:59 2025  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units   |
|--------------------------------|--------|------------|--------------|
| -----                          |        |            |              |
| System Monitoring Compounds    |        |            |              |
| 9) S ortho-Terphenyl (SURR)    | 11.676 | 5620598    | 37.873 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 75.75%       |
| 12) S 1-chlorooctadecane (S... | 13.111 | 4343627    | 38.436 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 76.87%       |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC072825AL\  
Data File : FC069547.D  
Signal(s) : FID1A.ch  
Acq On : 28 Jul 2025 15:18  
Operator : YP/AJ  
Sample : PB169021BL  
Misc :  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
FID\_C  
ClientSampleId :  
PB169021BL

Integration File: autoint1.e  
Quant Time: Jul 29 01:15:54 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 071525.M  
Quant Title : GC Extractables  
QLast Update : Tue Jul 15 12:35:59 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.676 min  
Delta R.T.: 0.010 min  
Response: 5620598  
Conc: 37.87 ug/ml

Instrument :  
FID\_C  
ClientSampleId :  
PB169021BL



#12 1-chlorooctadecane (SURR)

R.T.: 13.111 min  
Delta R.T.: 0.010 min  
Response: 4343627  
Conc: 38.44 ug/ml