

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC072922AL\
 Data File : FC060911.D
 Signal(s) : FID1A.ch
 Acq On : 30 Jul 2022 03:01
 Operator : YP/AJ
 Sample : N3964-03
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jul 30 06:00:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 072522.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 25 19:40:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	10.633	6596276	40.194 ug/ml
Spiked Amount 50.000		Recovery =	80.39%
12) S 1-chlorooctadecane (S...	11.998	5164056	38.899 ug/ml
Spiked Amount 50.000		Recovery =	77.80%
Target Compounds			
3) T A~Naphthalene (C11.7)	5.162	2394517	16.845 ug/ml
5) T A~2-methylnaphthalene...	6.204	395806	2.886 ug/ml
6) T n-Tetradecane (C14)	7.267f	228695	1.523 ug/ml
10) T n-Eicosane (C20)	11.536	221924	1.414 ug/ml
14) T n-Tetracosane (C24)	13.813	231289	1.515 ug/ml
16) T n-Octacosane (C28)	15.746	259990	1.814 ug/ml
20) T n-Hexatriacontane (C36)	18.940	203136	1.662 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC072922AL\
Data File : FC060911.D
Signal(s) : FID1A.ch
Acq On : 30 Jul 2022 03:01
Operator : YP/AJ
Sample : N3964-03
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jul 30 06:00:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 072522.M
Quant Title : GC Extractables
QLast Update : Mon Jul 25 19:40:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#3 A~Naphthalene (C11.7)


R.T.: 5.162 min
Delta R.T.: -0.022 min
Response: 2394517
Conc: 16.85 ug/ml

Instrument :
FID_C
ClientSampleId :




#5 A~2-methylnaphthalene (C12.89)

R.T.: 6.204 min
Delta R.T.: -0.010 min
Response: 395806
Conc: 2.89 ug/ml




#6 n-Tetradecane (C14)

R.T.: 7.267 min
Delta R.T.: -0.052 min
Response: 228695
Conc: 1.52 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 10.633 min
Delta R.T.: -0.002 min
Response: 6596276
Conc: 40.19 ug/ml



#10 n-Eicosane (C20)


R.T.: 11.536 min
Delta R.T.: -0.025 min
Response: 221924
Conc: 1.41 ug/ml

Instrument :
FID_C
ClientSampleId :




#12 1-chlorooctadecane (SURR)

R.T.: 11.998 min
Delta R.T.: 0.000 min
Response: 5164056
Conc: 38.90 ug/ml




#14 n-Tetracosane (C24)

R.T.: 13.813 min
Delta R.T.: 0.006 min
Response: 231289
Conc: 1.51 ug/ml



#16 n-Octacosane (C28)

R.T.: 15.746 min
Delta R.T.: 0.018 min
Response: 259990
Conc: 1.81 ug/ml



#20 n-Hexatriacontane (C36)

R.T.: 18.940 min
Delta R.T.: 0.048 min
Response: 203136
Conc: 1.66 ug/ml

Instrument :

FID_C

ClientSampleId :