

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC073024AL\
 Data File : FC066849.D
 Signal(s) : FID1A.ch
 Acq On : 30 Jul 2024 23:46
 Operator : YP/AJ
 Sample : P3396-07
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jul 31 03:29:26 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 071524.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 15 13:58:41 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.961	2723572	24.358 ug/ml
Spiked Amount 50.000		Recovery =	48.72%
Target Compounds			
4) T n-Dodecane (C12)	6.370	94732	0.632 ug/ml
6) T n-Tetradecane (C14)	8.165	71010	0.485 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC073024AL\
Data File : FC066849.D
Signal(s) : FID1A.ch
Acq On : 30 Jul 2024 23:46
Operator : YP/AJ
Sample : P3396-07
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jul 31 03:29:26 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 071524.M
Quant Title : GC Extractables
QLast Update : Mon Jul 15 13:58:41 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#4 n-Dodecane (C12)


R.T.: 6.370 min
Delta R.T.: -0.002 min
Response: 94732
Conc: 0.63 ug/ml

Instrument :
FID_C
ClientSampleId :




#6 n-Tetradecane (C14)

R.T.: 8.165 min
Delta R.T.: -0.002 min
Response: 71010
Conc: 0.49 ug/ml



#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min
Exp R.T. : 11.521 min
Response: 0
Conc: N.D.



#12 1-chlorooctadecane (SURR)

R.T.: 12.961 min
Delta R.T.: 0.000 min
Response: 2723572
Conc: 24.36 ug/ml