

rteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/02/2022
Supervised By :Ankita Jodhani 08/02/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC08012
Data File : FC060938.D
Signal(s) : FID1A.ch
Acq On : 02 Aug 2022 03:02
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 072522.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.625	2.589	2.735	BB	286193	2797144	90.28%	4.558%
2	3.619	3.592	3.729	BB	264883	2870722	92.65%	4.678%
3	5.160	5.125	5.359	BB	163437	2948751	95.17%	4.805%
4	5.565	5.535	5.665	BB	278412	2973285	95.96%	4.845%
5	6.183	6.149	6.380	BB	151314	2880090	92.95%	4.693%
6	7.314	7.274	7.407	BB	265233	2997150	96.73%	4.884%
7	8.875	8.834	8.954	BB	267578	3027996	97.73%	4.934%
8	10.281	10.244	10.352	BB	250431	3036800	98.01%	4.949%
9	10.631	10.584	10.735	BB	247803	3098382	100.00%	5.049%
10	11.560	11.520	11.615	BB	238751	2987301	96.41%	4.868%
11	11.996	11.957	12.059	BB	202872	2490905	80.39%	4.059%
12	12.158	12.120	12.219	BB	239764	2951614	95.26%	4.810%
13	12.730	12.684	12.780	BB	231638	2915392	94.09%	4.751%
14	13.806	13.767	13.864	BB	212031	2794792	90.20%	4.554%
15	14.803	14.740	14.857	BB	200804	2666717	86.07%	4.346%
16	15.730	15.657	15.772	BB	192368	2577532	83.19%	4.200%
17	16.596	16.522	16.635	BB	191740	2650297	85.54%	4.319%
18	17.408	17.335	17.457	BB	179620	2704554	87.29%	4.407%
19	18.172	18.060	18.227	BB	174425	2568672	82.90%	4.186%
20	18.893	18.710	18.935	BB	171438	2519912	81.33%	4.106%
21	19.576	19.337	19.620	BV	152325	2509766	81.00%	4.090%
22	20.224	20.012	20.279	BBA	151719	2397428	77.38%	3.907%
Sum of corrected areas:							61365202	

Aliphatic EPH 072522.M Tue Aug 02 11:12:42 2022