

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC080124AL\
 Data File : FC066885.D
 Signal(s) : FID1A.ch
 Acq On : 01 Aug 2024 13:46
 Operator : YP/AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Aug 02 02:56:16 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 071524.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 15 13:58:41 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	141.894	137.513 E3	3.1	96	0.00
2 T n-Decane (C10)	147.828	141.472 E3	4.3	96	0.00
3 T A~Naphthalene (C11.7)	159.695	151.941 E3	4.9	96	0.00
4 T n-Dodecane (C12)	149.917	141.799 E3	5.4	96	0.00
5 T A~2-methylnaphthalene (C12.	153.302	145.468 E3	5.1	95	0.00
6 T n-Tetradecane (C14)	146.312	138.405 E3	5.4	95	0.00
7 T n-Hexadecane (C16)	149.668	141.123 E3	5.7	95	0.00
8 T n-Octadecane (C18)	148.398	140.186 E3	5.5	95	0.00
9 S ortho-Terphenyl (SURR)	153.953	147.578 E3	4.1	97	0.00
10 T n-Eicosane (C20)	139.385	132.421 E3	5.0	96	0.00
11 T n-Heneicosane (C21)	136.210	129.062 E3	5.2	96	0.00
12 S 1-chlorooctadecane (SURR)	111.816	109.500 E3	2.1	98	0.00
13 T n-Docosane (C22)	134.320	127.967 E3	4.7	96	0.00
14 T n-Tetracosane (C24)	133.013	126.033 E3	5.2	96	0.00
15 T n-Hexacosane (C26)	130.117	122.875 E3	5.6	96	0.00
16 T n-Octacosane (C28)	126.154	121.343 E3	3.8	97	0.00
17 T n-Tricontane (C30)	128.594	124.094 E3	3.5	98	0.00
18 T n-Dotriacontane (C32)	125.893	122.106 E3	3.0	98	0.00
19 T n-Tetratriacontane (C34)	123.715	116.199 E3	6.1	94	0.00
20 T n-Hexatriacontane (C36)	116.325	109.280 E3	6.1	94	0.00
21 T n-Octatriacontane (C38)	111.107	105.998 E3	4.6	95	0.00
22 T n-Tetracontane (C40)	108.714	103.427 E3	4.9	95	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0