

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC080222AL\
 Data File : FC060960.D
 Signal(s) : FID1A.ch
 Acq On : 02 Aug 2022 22:10
 Operator : YP/AJ
 Sample : N3991-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 CL-02-080122-EPH-2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/03/2022
 Supervised By :Ankita Jodhani 08/03/2022

Integration File: autoint1.e
 Quant Time: Aug 03 06:18:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 072522.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 25 19:40:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	10.634	4797501	29.234 ug/mlm
Spiked Amount	50.000	Recovery	= 58.47%
12) S 1-chlorooctadecane (S...	12.001	3831598	28.862 ug/ml
Spiked Amount	50.000	Recovery	= 57.72%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC080222AL\
Data File : FC060960.D
Signal(s) : FID1A.ch
Acq On : 02 Aug 2022 22:10
Operator : YP/AJ
Sample : N3991-02
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
FID_C
ClientSampleId :
CL-02-080122-EPH-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/03/2022
Supervised By :Ankita Jodhani 08/03/2022

Integration File: autoint1.e
Quant Time: Aug 03 06:18:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 072522.M
Quant Title : GC Extractables
QLast Update : Mon Jul 25 19:40:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 10.634 min
Delta R.T.: 0.000 min
Response: 4797501
Conc: 29.23 ug/ml

Instrument :
FID_C
ClientSampleId :
CL-02-080122-EPH-2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/03/2022
Supervised By :Ankita Jodhani 08/03/2022



#12 1-chlorooctadecane (SURR)

R.T.: 12.001 min
Delta R.T.: 0.004 min
Response: 3831598
Conc: 28.86 ug/ml