

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC080322AL\
 Data File : FC060984.D
 Signal(s) : FID1A.ch
 Acq On : 04 Aug 2022 03:19
 Operator : YP/AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Aug 04 04:45:51 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 072522.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 25 19:40:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	10.631	2901508	17.680 ug/ml
Spiked Amount 50.000		Recovery =	35.36%
12) S 1-chlorooctadecane (S...	11.995	2323921	17.505 ug/ml
Spiked Amount 50.000		Recovery =	35.01%
Target Compounds			
1) T n-Nonane (C9)	2.628	2694952	19.135 ug/ml
2) T n-Decane (C10)	3.621	2767664	19.620 ug/ml
3) T A~Naphthalene (C11.7)	5.163	2837066	19.959 ug/ml
4) T n-Dodecane (C12)	5.567	2863744	19.321 ug/ml
5) T A~2-methylnaphthalene...	6.186	2616099	19.075 ug/ml
6) T n-Tetradecane (C14)	7.315	2876694	19.161 ug/ml
7) T n-Hexadecane (C16)	8.875	2888310	18.873 ug/ml
8) T n-Octadecane (C18)	10.282	2876087	18.418 ug/ml
10) T n-Eicosane (C20)	11.560	2810249	17.907 ug/ml
11) T n-Heneicosane (C21)	12.157	2762626	17.656 ug/ml
13) T n-Docosane (C22)	12.729	2722405	17.461 ug/ml
14) T n-Tetracosane (C24)	13.805	2611249	17.104 ug/ml
15) T n-Hexacosane (C26)	14.800	2518084	16.869 ug/ml
16) T n-Octacosane (C28)	15.727	2482956	17.328 ug/ml
17) T n-Tricontane (C30)	16.594	2648134	18.915 ug/ml
18) T n-Dotriacontane (C32)	17.405	2761178	20.459 ug/ml
19) T n-Tetratriacontane (C34)	18.170	2613390	20.837 ug/ml
20) T n-Hexatriacontane (C36)	18.891	2553015	20.886 ug/ml
21) T n-Octatriacontane (C38)	19.573	2480574	21.017 ug/ml
22) T n-Tetracontane (C40)	20.218	2369969	20.591 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC080322AL\
Data File : FC060984.D
Signal(s) : FID1A.ch
Acq On : 04 Aug 2022 03:19
Operator : YP/AJ
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
Quant Time: Aug 04 04:45:51 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 072522.M
Quant Title : GC Extractables
QLast Update : Mon Jul 25 19:40:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

