

rters

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC08122
Data File : FC061111.D
Signal(s) : FID1A.ch
Acq On : 13 Aug 2022 04:02
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Reviewed By :Yogesh Patel 08/15/2022
Supervised By :Ankita Jodhani 08/15/2022

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 072522.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.632	2.612	2.803	BB	244103	2842704	92.51%	4.571%
2	3.626	3.568	3.817	BB	225677	2866504	93.28%	4.609%
3	5.174	5.118	5.497	BB	106457	3014510	98.10%	4.847%
4	5.571	5.542	5.760	BB	242546	3029704	98.59%	4.872%
5	6.199	6.095	6.517	BB	96431	2905758	94.56%	4.672%
6	7.317	7.278	7.455	BB	244897	3016229	98.15%	4.850%
7	8.876	8.843	9.010	BB	250120	3069423	99.88%	4.936%
8	10.282	10.207	10.390	BB	252096	3072872	100.00%	4.941%
9	10.632	10.597	10.792	BB	239926	3072958	100.00%	4.941%
10	11.558	11.520	11.652	BB	245640	3012771	98.04%	4.844%
11	11.994	11.955	12.095	BB	201020	2471215	80.42%	3.974%
12	12.155	12.117	12.245	BB	226757	2963999	96.45%	4.766%
13	12.726	12.687	12.818	BB	224975	2902097	94.44%	4.667%
14	13.802	13.732	13.877	BB	207354	2761824	89.88%	4.441%
15	14.797	14.583	14.927	BB	194445	2884023	93.85%	4.637%
16	15.723	15.668	15.790	BB	185504	2539185	82.63%	4.083%
17	16.588	16.427	16.650	BB	173316	2615072	85.10%	4.205%
18	17.400	17.325	17.452	BB	184463	2678335	87.16%	4.307%
19	18.164	18.085	18.218	BB	186851	2629204	85.56%	4.228%
20	18.883	18.700	18.932	BB	169244	2639368	85.89%	4.244%
21	19.564	19.327	19.620	BB	162249	2625816	85.45%	4.222%
22	20.210	20.035	20.279	BBA	150311	2576289	83.84%	4.143%
Sum of corrected areas:							62189860	

Aliphatic EPH 072522.M Mon Aug 15 03:12:53 2022