

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC082222AL\
 Data File : FC061263.D
 Signal(s) : FID1A.ch
 Acq On : 24 Aug 2022 06:53
 Operator : YP/AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 24 07:16:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 082222.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 23 05:32:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.632	3346921	18.906 ug/ml
Spiked Amount 50.000		Recovery =	37.81%
12) S 1-chlorooctadecane (S...	13.063	2654438	18.953 ug/ml
Spiked Amount 50.000		Recovery =	37.91%
Target Compounds			
1) T n-Nonane (C9)	3.400	2982953	19.730 ug/ml
2) T n-Decane (C10)	4.467	2988342	19.694 ug/ml
3) T A~Naphthalene (C11.7)	6.052	3389004	19.578 ug/ml
4) T n-Dodecane (C12)	6.480	3021588	19.510 ug/ml
5) T A~2-methylnaphthalene...	7.105	3417773	19.410 ug/ml
6) T n-Tetradecane (C14)	8.273	3025991	19.346 ug/ml
7) T n-Hexadecane (C16)	9.872	3009608	19.214 ug/ml
8) T n-Octadecane (C18)	11.312	3020887	19.148 ug/ml
10) T n-Eicosane (C20)	12.618	2959505	19.037 ug/ml
11) T n-Heneicosane (C21)	13.229	2965810	18.961 ug/ml
13) T n-Docosane (C22)	13.815	2947322	18.993 ug/ml
14) T n-Tetracosane (C24)	14.915	2971651	19.209 ug/ml
15) T n-Hexacosane (C26)	15.935	2902810	19.287 ug/ml
16) T n-Octacosane (C28)	16.883	2834316	19.189 ug/ml
17) T n-Tricontane (C30)	17.769	2666576	19.306 ug/ml
18) T n-Dotriacontane (C32)	18.600	2122057	19.333 ug/mlm
19) T n-Tetratriacontane (C34)	19.381	1501043	19.447 ug/ml
20) T n-Hexatriacontane (C36)	20.119	1125460	19.934 ug/ml
21) T n-Octatriacontane (C38)	20.861	1000786	21.573 ug/ml
22) T n-Tetracontane (C40)	21.790	868264	20.701 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC082222AL\
Data File : FC061263.D
Signal(s) : FID1A.ch
Acq On : 24 Aug 2022 06:53
Operator : YP/AJ
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 24 07:16:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 082222.M
Quant Title : GC Extractables
QLast Update : Tue Aug 23 05:32:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

