

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC082522AL\
 Data File : FC061300.D
 Signal(s) : FID1A.ch
 Acq On : 25 Aug 2022 16:27
 Operator : YP/AJ
 Sample : N4328-03DL 5X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 1DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/26/2022
 Supervised By :Ankita Jodhani 08/26/2022

Integration File: sample.E
 Quant Time: Aug 25 18:16:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 082222.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 23 05:32:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.638	1334559	7.539 ug/mlm
Spiked Amount 50.000		Recovery =	15.08%
12) S 1-chlorooctadecane (S...	13.070	956598	6.830 ug/mlm
Spiked Amount 50.000		Recovery =	13.66%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC082522AL\
Data File : FC061300.D
Signal(s) : FID1A.ch
Acq On : 25 Aug 2022 16:27
Operator : YP/AJ
Sample : N4328-03DL 5X
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
FID_C
ClientSampleId :
1DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/26/2022
Supervised By :Ankita Jodhani 08/26/2022

Integration File: sample.E
Quant Time: Aug 25 18:16:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 082222.M
Quant Title : GC Extractables
QLast Update : Tue Aug 23 05:32:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.638 min
Delta R.T.: 0.018 min
Response: 1334559
Conc: 7.54 ug/ml

Instrument :
FID_C
ClientSampleId :
1DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/26/2022
Supervised By :Ankita Jodhani 08/26/2022



#12 1-chlorooctadecane (SURR)

R.T.: 13.070 min
Delta R.T.: 0.015 min
Response: 956598
Conc: 6.83 ug/ml m