

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC082818AL\
 Data File : FC036110.D
 Signal(s) : FID1A.CH
 Acq On : 28 Aug 2018 14:58
 Operator : AJ\SJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 29 00:40:55 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_C\Method\Alaphatic EPH 082118.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 21 16:09:01 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	89.822	98.057 E3	-9.2	106	0.00
2 T	n-Decane (C10)	90.009	99.454 E3	-10.5	106	0.00
3 T	A-Naphthalene (C11.7)	100.642	121.601 E3	-20.8	116	0.00
4 T	n-Dodecane (C12)	92.741	101.652 E3	-9.6	106	0.00
5 T	A-2-methylnaphthalene (C12)	100.690	122.516 E3	-21.7	116	0.00
6 T	n-Tetradecane (C14)	94.923	101.847 E3	-7.3	104	0.00
7 T	n-Hexadecane (C16)	95.190	101.057 E3	-6.2	103	0.00
8 T	n-Octadecane (C18)	98.235	101.848 E3	-3.7	102	0.00
9 S	ortho-Terphenyl (SURR)	111.504	128.856 E3	-15.6	112	0.00
10 T	n-Eicosane (C20)	96.635	101.541 E3	-5.1	103	0.00
11 T	n-Heneicosane (C21)	94.128	100.234 E3	-6.5	104	0.00
12 S	1-chlorooctadecane (SURR)	81.611	88.330 E3	-8.2	105	0.00
13 T	n-Docosane (C22)	93.128	99.311 E3	-6.6	104	0.00
14 T	n-Tetracosane (C24)	91.445	98.430 E3	-7.6	105	0.00
15 T	n-Hexacosane (C26)	89.698	96.364 E3	-7.4	105	0.00
16 T	n-Octacosane (C28)	90.997	94.567 E3	-3.9	103	0.00
17 T	n-Tricontane (C30)	96.223	91.779 E3	4.6	97	0.00
18 T	n-Dotriacontane (C32)	94.423	90.748 E3	3.9	99	0.00
19 T	n-Tetratriacontane (C34)	88.556	87.612 E3	1.1	101	0.00
20 T	n-Hexatriacontane (C36)	84.718	82.736 E3	2.3	99	0.00
21 T	n-Octatriacontane (C38)	83.389	77.672 E3	6.9	94	0.00
22 T	n-Tetracontane (C40)	83.970	73.368 E3	12.6	88	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0