

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC083019AL\  
 Data File : FC043267.D  
 Signal(s) : FID1A.CH  
 Acq On : 31 Aug 2019 15:44  
 Operator : DD\AJ  
 Sample : K4581-02DL 10X  
 Misc :  
 ALS Vial : 34 Sample Multiplier: 1

Instrument :  
 FID\_C  
 ClientSampleId :  
 T12-C1-C4-(0-3.5)DL

Integration File: autoint1.e  
 Quant Time: Aug 31 23:46:12 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Alaphatic EPH 082719.M  
 Quant Title : GC Extractables  
 QLast Update : Wed Aug 28 04:05:56 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response | Conc Units  |
|--------------------------------|--------|----------|-------------|
| -----                          |        |          |             |
| System Monitoring Compounds    |        |          |             |
| 9) S ortho-Terphenyl (SURR)    | 11.618 | 776889   | 5.426 ug/ml |
| Spiked Amount                  | 50.000 | Recovery | = 10.85%    |
| 12) S 1-chlorooctadecane (S... | 13.041 | 598032   | 5.474 ug/ml |
| Spiked Amount                  | 50.000 | Recovery | = 10.95%    |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC083019AL\  
Data File : FC043267.D  
Signal(s) : FID1A.CH  
Acq On : 31 Aug 2019 15:44  
Operator : DD\AJ  
Sample : K4581-02DL 10X  
Misc :  
ALS Vial : 34 Sample Multiplier: 1

Instrument :  
FID\_C  
ClientSampleId :  
T12-C1-C4-(0-3.5)DL

Integration File: autoint1.e  
Quant Time: Aug 31 23:46:12 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Alaphatic EPH 082719.M  
Quant Title : GC Extractables  
QLast Update : Wed Aug 28 04:05:56 2019  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.618 min  
Delta R.T.: -0.011 min  
Response: 776889  
Conc: 5.43 ug/ml

Instrument :  
FID\_C  
ClientSampleId :  
T12-C1-C4-(0-3.5)DL



#12 1-chlorooctadecane (SURR)

R.T.: 13.041 min  
Delta R.T.: -0.010 min  
Response: 598032  
Conc: 5.47 ug/ml