

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC083122AL\
 Data File : FC061393.D
 Signal(s) : FID1A.ch
 Acq On : 31 Aug 2022 22:50
 Operator : YP/AJ
 Sample : N4451-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 TP1

Integration File: autoint1.e
 Quant Time: Sep 01 05:18:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 082222.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 31 11:58:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.586	7580021	42.818 ug/ml
Spiked Amount 50.000		Recovery =	85.64%
12) S 1-chlorooctadecane (S...	13.022	6052229	43.215 ug/ml
Spiked Amount 50.000		Recovery =	86.43%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC083122AL\
Data File : FC061393.D
Signal(s) : FID1A.ch
Acq On : 31 Aug 2022 22:50
Operator : YP/AJ
Sample : N4451-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
TP1

Integration File: autoint1.e
Quant Time: Sep 01 05:18:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 082222.M
Quant Title : GC Extractables
QLast Update : Wed Aug 31 11:58:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.586 min
Delta R.T.: 0.008 min
Response: 7580021
Conc: 42.82 ug/ml

Instrument :
FID_C
ClientSampleId :
TP1



#12 1-chlorooctadecane (SURR)

R.T.: 13.022 min
Delta R.T.: 0.007 min
Response: 6052229
Conc: 43.21 ug/ml