

rteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report
Manual IntegrationsAPPROVED
Reviewed By :Yogesh Patel 09/15/2022
Supervised By :Ankita Jodhani 09/15/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC09142
Data File : FC061528.D
Signal(s) : FID1A.ch
Acq On : 15 Sep 2022 05:45
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 082222.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.334	3.300	3.385	BB	351755	2960517	88.63%	5.439%
2	4.401	4.368	4.453	BB	321349	2973422	89.02%	5.463%
3	5.986	5.930	6.018	BV	364016	3333271	99.79%	6.124%
4	6.418	6.352	6.468	BB	302630	2994761	89.66%	5.502%
5	7.043	6.978	7.072	BB	350293	3340217	100.00%	6.137%
6	8.216	8.158	8.260	BB	287022	2959044	88.59%	5.436%
7	9.819	9.752	9.863	BB	267860	2913576	87.23%	5.353%
8	11.262	11.215	11.305	BB	248922	2867479	85.85%	5.268%
9	11.578	11.512	11.633	BB	281190	3074892	92.06%	5.649%
10	12.572	12.497	12.608	BB	235487	2746358	82.22%	5.046%
11	13.017	12.942	13.067	BB	203001	2417008	72.36%	4.441%
12	13.184	13.117	13.225	BB	211613	2711824	81.19%	4.982%
13	13.771	13.678	13.823	BB	211880	2666179	79.82%	4.898%
14	14.873	14.807	14.913	BB	203344	2580254	77.25%	4.741%
15	15.894	15.828	15.940	BB	182715	2491928	74.60%	4.578%
16	16.844	16.442	16.880	BB	175286	2436315	72.94%	4.476%
17	17.732	17.677	17.773	BB	171818	2363168	70.75%	4.342%
18	18.564	18.527	18.602	M	136986	1961273	58.72%	3.603%
19	19.347	19.250	19.380	BV	98077	1510231	45.21%	2.775%
20	20.086	20.050	20.120	BV	75124	1114044	33.35%	2.047%
21	20.822	20.763	20.862	M	54104	1051538	31.48%	1.932%
22	21.734	21.683	21.790	M	38597	962494	28.82%	1.768%
Sum of corrected areas:						54429792		

Aliphatic EPH 082222.M Thu Sep 15 08:00:39 2022